

RULE OF THE CENTRAL BUREAU OF WEIGHTS AND MEASURES
ON THE PRESCRIPTION OF THE STANDARD AND PROVISION OR CONSTRUCTION OF
LABORATORY OR PLACE INCLUDING NECESSARY EQUIPMENT OF A PERSON APPLYING FOR
PERMISSION TO BE A SELF INSPECTOR OF A FUEL DISPENSER IN SERVICE STATION

B.E. 2551 (2008)

In order to prescribe that a producer of a fuel dispenser in service station who wishes to ask for permission to be an self inspector of a fuel dispenser produced by such applicant shall comply with the rule as prescribed in the Ministerial Regulation:

By virtue of section 6(5) of the Weights and Measures Act B.E.2542 and the provisions of clause 5(1) and clause 5(2) of the Ministerial Regulation prescribing the criteria for business operator relating to measurement B.E. 2544 issued under the Weights and Measures Act B.E. 2542 which is the Act containing certain provisions in relation to the restriction of rights and liberties of a person, in respect of which section 29 in conjunction with section 32, section 33, section 41 and section 43 of the Constitution of the Kingdom of Thailand so permit by the virtue of law, the Director-General of the Department of Internal Trade shall hereby issues the rules as follows:

Clause 1. This Rule is called the “Rule of the Central Bureau of Weights and Measures on the prescription of standard and provision or construction of a laboratory or place including necessary equipment of a person applying to be an self inspector of fuel dispenser in service station, B.E. 2551 (2008)”.

Clause 2. This Rule shall come into force as from the day following the date of its publication in the Government Gazette.

Clause 3. A producer of fuel dispenser in service station shall provide a stainless steel tank with scale with the smallest amount measured under the measurement system of fuel volume and capacity of 5 liters and 20 liters and 50 liters and shall be certified its

correctness from the Bureau of Weight and Measures every 1 year in order to use as the standard for accuracy of fuel dispenser in service station.

Clause 4. A producer of fuel dispenser in service station shall maintain the standard under clause 3 to be readily used at all times.

Clause 5. A producer of fuel dispenser in service station shall provide or construct a laboratory or place including necessary equipment for examining and certifying fuel dispenser in service station with the following characteristics:

(1) being a laboratory or place which is stable and permanent, strong, safe and correct in accordance with the plan and details submitted to the Director-General and located in the place where there are convenient routes between the location of the central Bureau or subsidiary office;

(2) having an appropriate area for operation to install a standard for inspection of fuel dispenser in service station;

(3) having a pump for fuel which ensures that the flow rate is not less than 50 liter per minute for the performance of inspection of the accuracy of fuel dispenser in service station;

(4) having an electricity system which provides sufficient illumination;

(5) having necessary and sufficient equipment and mechanic tools.

Clause 6. The Director-General of the Department of Internal Trade shall have change and control of the execution of this Rule.

Given on the 28th of March B.E. 2551 (2008)

Yanyong Puangraj

Director-General of the Department of Internal Trade