The report of the test for the protective ability against a magnetic field

Application number	
Trademark/model	Day/Month/Year of the test
Test machine number	. The number of water volume metres Units

Order	I.D.	Qn cubic metre/hr			Qt cubic metre/hr			Qmin metre/hr				Test result		
	number	Vol. tested litres			Vol. tested litres			Vol. tested litres						
	(magnet	Inflow pressure kpascal			Inflow	Inflow pressure kpascal			Inflow pressure kpascal					
	installing	Outflow pressure			Outflow pressure			Outflow pressure						
	position)	kpascal			kpascal			kpascal						
		Met	re	Margin	of	Metre Margin of			Met	Margin	of			
		error +/-		-/-	error +/-			error +/-			-/-			
				2%				2%				5%		
		Before	After	Diff.	%	Before	After	Diff.	%	Before	After	Diff.	%	
				margin				margin				margin		
1														Pass or fail
2			1											Pass or fail
3														Pass or fail
4		N												Pass or fail
5														Pass or fail
6														Pass or fail
7														Pass or fail
8														Pass or fail
9														Pass or fail
10														Pass or fail
11														Pass or fail
12														Pass or fail
13														Pass or fail
14														Pass or fail
15														Pass or fail
16														Pass or fail

17							Pass or fail
18							Pass or fail
19							Pass or fail
20							Pass or fail
21							Pass or fail
22							Pass or fail
23							Pass or fail
24				·			Pass or fail
25							Pass or fail

The criteria:	\bigcirc	Pass		Fai					
In case of sample	testing: if t	the result c	of the test sho	ws that o	one or more or all of the units of				
the water volume metre fails to pass the test, the rest of the units that are not previously									
collected as samples shall be tested.									
Signature		(······) The license holder or an agent				